

Mannheim, July 18, 2005

To whom it may concern,

Self Declaration


Blood Glucose Meters ACCU-CHEK® Compact, ACCU-CHEK® Compact Plus, ACCU-CHEK® Go and ACCU-CHEK® Active supplied by Roche Diagnostics GmbH

Herewith we certify that the above mentioned meters for the purpose of monitoring glucose concentration in human blood are tested according to Environmental Conditions and Test Procedures for Airborne Equipment (RTCA/DO-160D) Section 21 Emission of Radio Frequency Energy.

The meters fulfill the requirements according to this standard, category M.

The meters do not effect the performance of aircrafts.

Roche Diagnostics GmbH
i. V.



Rainer Ringelstein
Head of Quality
Diabetes Care Systems

i. V.



Dr. Günter Jungwirt
Head of Complaint Handling
CAPA & Trending
Diabetes Care Systems

Self Declaration-s

Roche Diagnostics GmbH

Roche Diabetes Care
Sandhofer Strasse 116
D-68305 Mannheim
Telefon +49-621-7590
Telefax +49-621-7592890

Registergericht Mannheim
HRB 3962
Aufsichtsrat:
Dr. Franz B. Humer, Vorsitzender

Geschäftsführung:
Dr. Jürgen Schwiezer, Vorsitzender
Dr. Manfred Baier, Staffan Ek,
Dr. Volker Pfahlert,
Burkhard G. Piper,
Peter-Claus Schiller,
Prof. Dr. Dr. Klaus Strein