



16 June 2017

To Whom It May Concern:

Subject: Self Declaration

Blood Glucose Meters Accu-Chek Guide supplied by Roche Diabetes Care.

Herewith we certify that the above meters for the purpose of monitoring glucose concentration in human blood are tested according to Environmental Conditions and Test Procedures for Airborne Equipment (RTCA/DO-160E) Section 21 Emission of Radio Frequency Energy.

The meters fulfill the requirements according to this standard, Category M.

The meters do not affect the performance of aircrafts.

Nigel Surridge

Vice President R&D

Roche Diabetes Care

Greg Smith

Vice President of Regulatory

Roche Diabetes Care