

To whom it may concern

28 July 2011

Self Declaration

Concerning the Blood Glucose Meters
ACCU-CHEK® Compact Plus, ACCU-CHEK® Compact Plus [Model GT],
ACCU-CHEK® Active [Model GC],
ACCU-CHEK® Go, ACCU-CHEK® Go [Model GS] and
ACCU-CHEK® Mobile
supplied by Roche Diagnostics GmbH

Herewith we certify that the above mentioned meters for the purpose of monitoring glucose concentration in human blood are tested according to Environmental Conditions and Test Procedures for Airborne Equipment (RTCA/DO-160) Section 21 Emission of Radio Frequency Energy.

The meters fulfill the requirements according to this standard, category M.

The meters do not effect the performance of aircrafts.


Sincerely,

Roche Diagnostics GmbH
i.V. / on behalf of the company

i.V. / on behalf of the company



Dr. Malte Rutschmann
Head of Global Regulatory Affairs



Ringelstein
Head of Quality Diabetes Care Systems

Roche Diagnostics GmbH

Diagnostics Division

DCQAM
Bldg/Room 161/E..261

Tel. 0621-759-2263
Fax 0621-759-3801

Roche Diagnostics GmbH; Sandhofer Strasse 116; D-68305 Mannheim; Telefon +49 621 759 0; Telefax +49 621 759 2890

Sitz der Gesellschaft: Mannheim - Registergericht: AG Mannheim HRB 3962 - Geschäftsführung: Thomas Schmid, Sprecher; Edgar Vieth - Aufsichtsratsvorsitzender: Dr. Severin Schwan