



13 January 2012

To Whom It May Concern:

Subject: SELF DECLARATION

Blood Glucose Meters ACCU-CHEK[®] Aviva Nano supplied by Roche Diagnostics.

Herewith we certify that the above meters for the purpose of monitoring glucose concentration in human blood are tested according to Environmental Conditions and Test Procedures for Airborne Equipment (RTCA/DO-160E) Section 21 Emission of Radio Frequency Energy.

The meters fulfill the requirements according to this standard, category M.

The meters do not affect the performance of aircrafts.

Karen Roberts
Director of Design Quality
Diabetes Care Systems

1-13-2012

Gregory Smith
Director of Quality
Diabetes Care Systems

1/13/2012